TECHNICAL DATA MODEL DLM-1000

1. SENSOR

Model DLM-1000 "B" UV Sensor	320-400 nm (µW/cm²)
Model DLM-1000 "C" Visible Sensor	380-700 nm (foot-candles)
Model DLM-1000 "L" Visible Sensor	380-700 nm (Lux)
(Photometric Response)	

Calibration: Sensors are calibrated using line sources, or if more appropriate for the application, wide bandwidth sources. Calibration is within 5% of a working standard, traceable to the NATIONAL INSTITUTE OF STANDARDS and TECHNOLOGY, formerly NBS. Temperature coefficient±0.25%/°C.

2. READOUT UNIT

Display 13mm	(.0.5 in.) constant fluorescing
	Liquid Crystal Display (LCD)
Sampling time	0.4 seconds

Electrical Specifications:

Range	Resolution	Accuracy 25° ± 5° C	<u>Units</u>
0-1999	1	± 2% (+2 digits)	FC or Lux
0-19990	10	± 2% (+2 digits)	µW/cm²

Measurement Units

Spectral Range	<u>Model</u>	<u>Units</u>
320 - 400 nm	DLM-1000 "B" UV sensor	μW/cm²
380 - 700 nm	DLM-1000 "C" visible sensor	FC
380 - 700 nm	DLM-1000 "L" visible sensor	Lux

Operation

- 1. Slide the "POWER" switch to the "ON" position.
- Turn the lamp on or apply power to the light source being measured.
- Place the sensor where the radiance/luminance is to be measured.
- 4. The value of the radiance/luminance will be indicated on the display. If at any time only a "1" appears in the left hand display position, then the measurement is out of range. Indicated values are as follows:

Lux or foot-candles 1 x the display reading Microwatts/cm² 10 x the display reading

For further information please contact:



1431 W. Second Street " Pomona, CA 91766

3. POWER REQUIREMENT

Battery Operation	One 9-volt battery provided
Current Drain	2 mA (approx)
Battery Life	150 - 200 hours

4. TEMPERATURE RANGE

Operation	0° - 50° C (32° - 122°F)
Storage	10°C - 50° C (14° - 122°F)

5. HUMIDITY RANGE

0 to 100% R.H. non-condensing

6. DIMENSIONS

Readout Unit	. 108mm x 73mm x 23mm
	(4.3in. x 2.9in. x 0.9in.)
Sensor Head	. 70mm x 52 mm x 17 mm
	(2.8in. x 2.0 in. x 0.7 in.)

7. WEIGHT

Readout Unit	160 g (0.36 lb.)
Sensor	90 g (0.20 lb.)

8. USAGE

"B" UV sensor is used to measure blacklight output UVA.

"C" visible sensor is used to measure incandescent light output and to test for visible light which may be emitted by a Blacklight. This sensor is calibrated to read in foot-candles.

"L" visible sensor is the same as the "C sensor, except it is calibrated to read in Lux. (Optional)

Calibration

The visible and UV sensor radiometer should be returned to the factory every six months for recalibration or a routine check-up. The recommended six month calibration is based upon normal usage of intermittent readings. If the sensors are used continuously, recalibration should be performed more often. Visible sensors are calibrated to accurately read incandescent light sources, unless a fluorescent light calibration is requested.

Represented by:

BERG ENGINEERING & SALES COMPANY, INC.

3893 Industrial Avenue • Rolling Meadows, IL • 60008 Phone: (847) 577-3980 • Fax: (847) 577-0474 www.BergEng.com • Info@BergEng.com

Your Complete Source for Testing Equipment Since 1969!